

Product: 2019-nCoV CDC EUA Kit, 500 rxn

Product number: 10006606

Batch number: 0000508785

Expiration date: 2022-MAR-10

Component	Component Part Number	Component Batch Number
2019-nCoV_N1 Aliquot, 500 rxn	10006600	0000508024
2019-nCoV_N2 Aliquot, 500 rxn	10006601	0000508162
2019-nCoV_N3 Aliquot, 500 rxn	10006602	0000508263
RNAse P Aliquot, 500 rxn	10006603	0000508398

Quality Control Parameters		
Test	Specification	Results
ESI	±2.8 Da of target mass of each primer and probe	Pass
Purity by CE	All primers ≥85% full length product	Pass
Purity by CE	All probes ≥80% full-length product	Pass

Storage: Store 2019-nCoV CDC EUA Kit, 500 rxn in a sealed container at –20°C.

Verified by: Natalie Walter

Quality release date: 2020-MAR-20

This product has not been clinically validated and is not designed or intended for use in clinical application. The product identified herein is manufactured under IDT ISO certified quality systems in compliance with GMP and in accordance with approved IDT analytical specifications. This certifies that the product has been reviewed and all analytical specifications have been determined to be acceptable. This document was produced electronically and is valid without signature.

For Research Use Only. Not for use in diagnostic procedures.

IDT is a trademark of Integrated DNA Technologies, Inc.