

Product: Resuspension Buffer (IDTE) pH 8.0, 1 L

Product number: 1004711

Batch number: 0000431329

Expiration date: 2021-FEB-20

Test	Specification	Results
Visual description	Clear, colorless liquid	Pass
Determination of particulates	No particulates	Pass
Determination of pH	7.8–8.2	Pass
Endotoxin contamination	<0.25 EU/mL	Pass
RNase contamination	<1 ng/mL RNase A <12.5 ng/mL RNase T1 <70 ng/mL RNase I	Pass

Storage: Store Resuspension Buffer (IDTE) pH 8.0 in a sealed container in ambient conditions.

Verified by: Ekaterina Carlson

Quality release date: 2019-FEB-22

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

IDT is a trademark of Integrated DNA Technologies, Inc.