

**Product:** Library Preparation Component Box, 24 rxn

**Product number:** 1056727

**Batch number:** 0000424136

**Omixon® batch number:** R2/017

**Expiration date:** 2020-MAY-07

**Manufacturing date:** 2019-JAN-16

Component	Component PN	Batch number	Test	Specification	Results
Fragmentation enzyme (A) 24	1056716	3161214	Volume	70 µL ±10%	Pass
Fragmentation buffer (B) 24	1056715	7371214	Volume	70 µL ±10%	Pass
End repair enzyme (C) 24	1056714	3141214	Volume	41 µL ±10%	Pass
End repair buffer (D) 24	1056713	891214	Volume	82 µL ±10%	Pass
Ligation enzyme (E) 24	1056712	0551214	Volume	81 µL ±10%	Pass
Ligation buffer (F) 24 [x2]	1056711	0541214	Volume	900 µL ±10%	Pass

**Storage:** Store Library Preparation Component Box in a sealed container at –20°C.

**Verified by:** Natalie Walter

**Quality release date:** 2019-JAN-16

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

IDT is a trademark of Integrated DNA Technologies, Inc.

**Product:** Library Preparation Component Box, 24 rxn

**Product number:** 1056727

Part number	Lot number	COA release date
1056727	0000424136	2019-JAN-16

**Verified by:** Natalie Walter

**Quality release date:** 2019-JAN-16

Integrated DNA Technologies hereby declares that Omixon® product Library Preparation Component Box, 24 rxn, batch: 0000424136 has been manufactured in full compliance with Omixon-defined specifications contained in the master specification document and that it has passed Integrated DNA Technologies QC test as documented in the associated Certificate of Analysis (COA).

IDT is a trademark of Integrated DNA Technologies, Inc.