INTEGRATED DNA TECHNOLOGIES

certificate of analysis

Product: xGen Universal Blockers—10 bp TS Mix, 4 x 96 rxn

Product number: 1081102

Batch number: 0000420458

Expiration date: 2021-NOV-07

Test	Specification	Results
Purity by analytical RP HPLC	>75% full-length product	Pass
Concentration	±15% of theoretical concentration	Pass
Molecular weight by ESI	All species present ±0.02% of theoretical molecular weight for each	Pass

Storage: Store xGen Universal Blockers—10 bp TS Mix in a sealed container at –20°C.

Verified by: Natalie Walter
Quality release date: 2018-DEC-21

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

xGen and IDT are trademarks of Integrated DNA Technologies, Inc.