

**Product:** Resuspension Buffer (IDTE) pH 8.0 Pack, 4 x 1 L

**Product number:** 11-05-02-04

**Batch number:** 0000424552

**Expiration date:** 2021-JAN-09

Test	Specification	Results
Visual description	Clear, colorless liquid	Pass
Determination of particulates	No particulates	Pass
Determination of pH	7.8–8.2	Pass
Endotoxin contamination	<0.25 EU/mL	Pass
RNase contamination	<1 ng/mL RNase A <12.5 ng/mL RNase T1 <70 ng/mL RNase I	Pass

**Storage:** Store Resuspension Buffer (IDTE) pH 8.0 Pack in a sealed container in ambient conditions.

**Verified by:** Natalie Walter

**Quality release date:** 2019-JAN-21

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.