

Product: Nuclease-Free Water Pack, 4 x 1 L

Product number: 11-05-02-02

Batch number: 0000422814

Expiration date: 2021-JAN-03

Test	Specification	Results
Visual description	Clear, colorless liquid	Pass
Determination of particulates	No particulates	Pass
Conductivity	$\leq 2 \mu\text{S/cm}$	Pass
Endotoxin contamination	$< 0.25 \text{ EU/mL}$	Pass
DNase contamination	$< 0.005 \text{ U DNase I}$	Pass
RNase contamination	$< 1 \text{ ng/mL RNase A}$ $< 12.5 \text{ ng/mL RNase T1}$ $< 70 \text{ ng/mL RNase I}$	Pass

Storage: Store Nuclease-Free Water Packs in a sealed container in ambient conditions.

Note: This product was not treated with DEPC to remove potential nucleases.

Verified by: Wade Beagle

Quality release date: 2019-JAN-11

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

IDT is a trademark of Integrated DNA Technologies, Inc.