INTEGRATED DNA TECHNOLOGIES

certificate of analysis

Product: Nuclease-Free Water Pack, 4 x 1 L

Product number: 11-05-02-02

Batch number: 0000440117

Expiration date: 2021-FEB-09

Test	Specification	Results
Visual description	Clear, colorless liquid	Pass
Determination of particulates	No particulates	Pass
Conductivity	≤2 µS/cm	Pass
Endotoxin contamination	<0.25 EU/mL	Pass
DNase contamination	<0.005 U DNase I	Pass
RNase contamination	<1 ng/mL RNase A <12.5 ng/mL RNase T1 <70 ng/mL RNase I	Pass

Storage: Store Nuclease-Free Water Packs in a sealed container in ambient conditions.

Note: This product was not treated with DEPC to remove potential nucleases.

Verified by: Wade Beagle
Quality release date: 2019-APR-05

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

IDT is a trademark of Integrated DNA Technologies, Inc.