

Product: AspID Kit

Product number: 1079106

Lot number: 217065412

Expiration date: Refer to the expiration date printed on individual vial labels.

Test*	Specification	Results
AspID positive ctrl		
Functional performance	Pass	Pass
Volume	>250 µL	Pass
OLM AspID IEC DNA		
Functional performance	Pass	Pass
Volume	>250 µL	Pass
AspID no template control		
Functional performance	Pass	Pass
PT gene expression master mix		
Functional performance	Pass	Pass
Volume	>500 µL	Pass
Purified water nuclease free		
Functional performance	Pass	Pass
Volume	>100 µL	Pass
AspID primer / probe mix		
Functional performance	Pass	Pass
Volume	>100 µL	Pass

* Cross contamination of each component is assessed through a series of qPCR assays.

Storage: Store AspID Kit in a sealed container at –20°C.

Verified by: Jacob Petrzela

Quality release date: 2020-JAN-07

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

IDT is a trademark of Integrated DNA Technologies, Inc.