

Product: DNaseAlert QC Kit

Product number: 1075376

Batch number: 282112097

Expiration date: 2020-JAN-04

Test	Specification	Results
Functional testing	Signal-to-noise ratio >50	Pass
Quantification	2 nmol ±15%	Pass

Packaging: This batch contains 100-2 nmol tubes packaged in a double-bagged amber baggie.

Storage: Store DNaseAlert QC Kit in a sealed container at -20°C, protected from light.

Verified by: Wade Beagle

Quality release date: 2019-JAN-08

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

DNaseAlert and IDT are trademarks of Integrated DNA Technologies, Inc.