

Product: Nuclease-Free Water, 10 x 2 mL

Product number: 11-04-02-01

Batch number: 0000482327

Expiration date: 2021-OCT-27

Test	Specification	Results
Visual description	Clear, colorless liquid	Pass
Determination of particulates	No particulates	Pass
Conductivity	$\leq 2 \mu\text{S}/\text{cm}$	Pass
Endotoxin contamination	$< 0.25 \text{ EU}/\text{mL}$	Pass
DNase contamination	$< 0.005 \text{ U DNase I}$	Pass
RNase contamination	$< 1 \text{ ng}/\text{mL RNase A}$ $< 12.5 \text{ ng}/\text{mL RNase T1}$ $< 70 \text{ ng}/\text{mL RNase I}$	Pass

Storage: Store Nuclease-Free Water in a sealed container at 2–8°C.

Note: This product was not treated with DEPC to remove potential nucleases.

Verified by: Jacob Petrzela

Quality release date: 2019-NOV-12

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.