

Product: xGen Hybridization and Wash Kit, 96 rxn

Catalog number: 1080584

Product number (box 1 of 2): 1080585

Batch number (box 1 of 2): 0000474118

Expiration date: 2021-MAR-07

Component	Component part #	Batch number	Test	Specification	Results
2X Hybridization buffer	1072277	0000472138	pH	8.0 ±0.1	Pass
Hybridization buffer enhancer	1080587	0000472734	No testing required		
Bead wash buffer	1072279	0000472147	pH	7.5 ±0.1	Pass
10X Stringent wash buffer	1072276	0000472176	pH	6.5 ±0.1	Pass
10X Wash buffer I	1072273	0000472141	pH	6.8 ±0.1	Pass
10X Wash buffer II	1072274	0000472145	pH	5.9 ±0.1	Pass
10X Wash buffer III	1072275	0000472174	pH	5.6 ±0.1	Pass
Human Cot DNA	1080685	0000473846	Concentration*	1.0–1.2 mg/mL	Pass
			Purity (A ₂₆₀ /A ₂₈₀)*	1.8–2.1	Pass

* Results reported by vendor.

Test	Specification	Result
Functional performance: On-target read percentage (+150 bp flank) [†]	≥80.0%	85.9%

[†] Functional performance is assessed in a capture experiment using 500 ng of gDNA library input, the xGen Acute Myeloid Leukemia Cancer Panel, a qualified batch of Human Cot DNA, and streptavidin beads.

Storage: Store xGen Hybridization and Wash Kit in a sealed container at –20°C.

Verified by: Wade Beagle

Quality release date: 2019-SEP-27

IDT verifies that the information contained herein is true and correct to the best of our knowledge. This document was produced electronically and is valid without signature.

For Research Use Only.

xGen is a trademark of Integrated DNA Technologies, Inc.