

Notices

Limitations of use

For research use only. Not for use in diagnostic procedures. Unless otherwise agreed to in writing, IDT does not intend these products to be used in clinical applications and does not warrant their fitness or suitability for any clinical diagnostic use. Purchaser is solely responsible for all decisions regarding the use of these products and any associated regulatory or legal obligations.

Safety data sheets pertaining to this product are available upon request.

Safety Notices



Reminder symbols call attention to minor details that may be easily overlooked and compromise the procedure resulting in decreased assay performance.



Caution symbols denote critical steps in the procedure where risk of protocol failure or damage to the product itself could occur if not carefully observed.



Stop symbols indicate where this procedure may be safely suspended and resumed at a later time without risk of compromised assay performance. Make note of these steps and plan your workflow accordingly.

Contents

Overview.....	3
Before getting started.....	8
Protocol.....	11

Additional resources

View videos and additional resources for Archer products at <http://archerdx.com/videos>.

Technical support

Visit <https://support.archerdx.com> for a list of helpful answers to frequently asked questions or contact us directly at archer-tech@idtdna.com.

Overview

Intended use

The Archer Immunoverse-HS TCR protocol is intended for research applications in conjunction with Archer reagent kits and corresponding target-enrichment panels to produce high-complexity libraries for use with Illumina next-generation sequencing (NGS) platforms.

Immunoverse sequencing data produced by this method should be processed using Archer[™] Analysis software—a complete bioinformatics suite that leverages Anchored Multiplex PCR (AMP[™]) chemistry to identify unique sequence fragments, thus enabling error correction, read deduplication, and ultimately high-confidence alignment and mutation calling. Archer Analysis takes demultiplexed FASTQ files straight from the sequencer as input and produces both high-level and detailed mutation reporting, as well as raw text and BAM outputs for full transparency of the pipeline.

Test principle

Anchored Multiplex PCR (AMP) is a rapid and scalable method to generate target-enriched libraries for NGS. AMP technology can be used for applications in targeted RNA sequencing, genomic DNA sequencing, and genotyping applications to generate a sequencing library in a matter of hours. Designed for low nucleic acid input, this process delivers robust performance across a variety of sample types.

AMP utilizes unidirectional gene-specific primers (GSPs) that enrich for both known and unknown mutations. Adapters that contain both molecular barcodes and sample indices permit quantitative multiplex data analysis, read deduplication, and accurate clonotype calling.

The Archer Analysis software utilizes these molecular barcodes (MBCs) for duplicate read binning, error correction and read deduplication to support quantitative multiplex data analysis and confident clonotype detection. Analysis reports both sequencing metrics and number of unique observations supporting called clonotypes.


Modular assay format Archer library preparation reagents include:

- Archer Immunoverse-HS TCR reagents in lyophilized format for each step of library preparation.
- GSPs that target panel-specific regions of interest during PCR amplification.
- Archer MBC adapters are proprietary adapters that tag each unique molecule with a barcode and common region prior to amplification.

Modular assay format

Archer kits include library preparation reagents and assay-specific liquid primers, which are used in conjunction with Archer MBC adapters to construct sequencing-ready libraries from total nucleic acid (TNA) or RNA samples.

Workflow



Immunoverse-HS TCR Reagents for Illumina (dSK0159)

Store at 2°C to 8°C - Allow pouches to reach room temperature before opening.

Materials Supplied		
Description	Part Number	Quantity
<i>TCR Specific RT Priming</i>	SA0201	1 pouch (8 reactions/one 8-tube strip)
<i>First Strand cDNA Synthesis</i>	SA0002	
<i>Second Strand cDNA Synthesis</i>	SA0003	
<i>End Repair</i>	SA0204	
<i>Ligation Step 1, 2.0</i>	SA0202	
<i>Ligation Step 2</i>	SA0197	
<i>First PCR (Illumina-HS)</i>	SA0359	
<i>Second PCR Reactions 1 thru 8 (Illumina- HS)</i>	SA0360	
<i>Ligation Cleanup Beads</i>	SA0210	1 tube
<i>Ligation Cleanup Buffer</i>	SA0209	(sufficient for processing 8 samples)

Additional components

Store at -30°C to -10°C

Materials Supplied		
Description	Part Number	Quantity
<i>Immunoverse GSP</i>	<i>Varies</i>	<i>Varies</i>



Materials required, but not supplied have been optimized for proper application of the Archer Library preparation. Use of other materials have not been tested by IDT.

Materials Required, but not Supplied		
Description	Supplier	Part Number
<i>Archer MBC adapters for Illumina</i>	<i>Archer</i>	<i>Varies</i>
<i>AMPure[®] XP beads</i>	<i>Beckman Coulter</i>	<i>A63880</i>
<i>1 M Tris-HCl, pH 8.0 (molecular biology grade)</i>	<i>Various</i>	<i>-</i>
<i>Ultrapure water (molecular biology grade)</i>	<i>Various</i>	<i>-</i>
<i>200 mM Tris-HCl, pH 7.0 (for sequencing)</i>	<i>Various</i>	<i>-</i>
<i>100% ethanol (ACS grade)</i>	<i>Various</i>	<i>-</i>
<i>Concentrated NaOH solution (ACS grade)</i>	<i>Various</i>	<i>-</i>
<i>RNase AWAY[™]</i>	<i>Thermo Fisher Scientific</i>	<i>7003</i>
<i>KAPA Universal Library Quantification Kit</i>	<i>KAPA Biosystems</i>	<i>KK4824</i>
<i>MiSeq[®] or NextSeq[®] Reagent Kit (300 cycle minimum)</i>	<i>Illumina</i>	<i>-</i>
<i>PhiX Control v3</i>	<i>Illumina</i>	<i>FC-110-3001</i>
<i>Standard PCR thermal cycler</i>	<i>Various</i>	<i>-</i>
<i>Real-Time PCR thermal cycler</i>	<i>Various</i>	<i>-</i>
<i>qPCR tubes</i>	<i>Various</i>	<i>-</i>
<i>0.2 mL PCR tubes</i>	<i>Various</i>	<i>-</i>
<i>DynaMag[™]-96 Side Magnet</i>	<i>Thermo Fisher Scientific</i>	<i>12331D</i>
<i>Microcentrifuge</i>	<i>Various</i>	<i>-</i>
<i>Plate centrifuge</i>	<i>Various</i>	<i>-</i>
<i>Pipettes (P10, P20, P200 and P1000)</i>	<i>Pipetman or equivalent</i>	<i>-</i>

<i>Sterile, nuclease-free aerosol barrier pipette tips</i>	<i>Various</i>	-
<i>Vortex mixer</i>	<i>Various</i>	-
<i>PCR tube cooling block</i>	<i>Various</i>	-
<i>Gloves</i>	<i>Various</i>	-
<i>Qubit[®] Fluorometer (3.0 or higher)</i>	<i>Thermo Fisher Scientific</i>	<i>Q33216</i>
<i>Qubit RNA HS Assay Kit</i>	<i>Thermo Fisher Scientific</i>	<i>Q32852</i>

Before getting started

Important precautions

- **Read through the entire protocol before starting your library preparation.**
- Take note of safe stopping points throughout the protocol where samples can be safely frozen (-30°C to -10°C) to plan your workflow.
- Use good laboratory practices to prevent contamination of samples by PCR products.
- Use nuclease-free PCR tubes, microcentrifuge tubes, and aerosol-barrier pipette tips.
- Wipe down workstation and pipettes with nuclease and nucleic acid cleaning products (e.g., RNase AWAY, Thermo Fisher Scientific).
- Verify that the thermal cycler used for library preparation is in good working order and currently calibrated according to manufacturer specifications. When using a thermal cycler with a 100% ramp rate >6°C/sec, set the ramp rate to go no higher than 6°C/sec for First PCR and Second PCR.
- Reaction cleanup with AMPure XP beads (Beckman Coulter) is performed at room temperature (20°C to 25°C) and is used repeatedly throughout the workflow. Ensure that AMPure XP beads are equilibrated to room temperature and fully resuspended by vortexing until homogenous in both color and appearance prior to drawing out material for **each** use.

Working with lyophilized reagents


- Archer reagents are provided as individually lyophilized reaction pellets in 0.2 mL PCR tube strips.
- Allow pouches to reach room temperature (20°C to 25°C) before opening to prevent moisture condensation on tubes.
- Always centrifuge tubes briefly before opening to pull contents down.
- Detach the required number of reaction tubes and return any unused portion to the pouch with desiccant packet, reseal and store at 2°C to 8°C. **It is recommended to use the remaining reactions within 4 weeks after opening.**
 - For MBC adapters and Second PCR tubes remember to label prior to returning unused portions to storage.
- Dissolve, mix and spin down:
 - **Never touch the lysosphere with the pipette tip.**

- Add sample/reagents to pellets in tubes while on ice.
- Allow at least 5 seconds for pellets to dissolve.
- Pipette up and down 8 times to mix after the lysosphere has dissolved.
- Briefly centrifuge and return to ice before proceeding.

Input nucleic acid

- Input nucleic acid (TNA or RNA) in EDTA-free buffer (pH 7-8) or ultrapure water is the optimal starting template for AMP library preparation. Do NOT use EDTA-containing buffers.
- Use the maximum allowable input mass (ng) whenever possible. The minimum recommended input mass for Immunoverse assays is dependent on experiment type, tissue type, and tissue quality:
 - 20 ng of RNA for screening of high abundance clones from high-quality PBMC samples
 - 400 ng of RNA for deep sequencing applications from high-quality, high-complexity PBMC samples
 - 400 ng of RNA from fresh-frozen tissue samples
 - > 400 ng of RNA for FFPE input for TIL analysis
- The rarefaction analysis below derived from varying amounts of high-quality input from a DNase-treated PBL sample illustrates the relationship between input amounts, sequencing depths and clonotype count:

Rarefaction Analysis



- Contact technical support (archer-tech@idtdna.com) for commercially available extraction kit recommendations.

Reagents to prepare before starting

- Make at least **10 mL** fresh 10 mM Tris-HCl, pH 8.0 from 1M Tris-HCl, pH 8.0 and ultrapure water.
 - Mix **100 µL** 1M Tris-HCl, pH 8.0 with **9,900 µL** ultrapure water.
 - 10 mM Tris-HCl, pH 8.0 can be used for up to one week after mixing.
- Make at least **50 mL** fresh 70% ethanol from 100% ethanol and ultrapure water.
 - Add **35 mL** 100% ethanol to **15 mL** ultrapure water.
 - Tightly close the cap to minimize evaporation when not in use.
 - 70% ethanol is appropriate for use for up to one week after mixing.
- Make at least **1 mL** fresh 5 mM NaOH from concentrated NaOH and ultrapure water on the day of use.
 - If working from 1M NaOH, add **5 µL** of 1M NaOH to **995 µL** of ultrapure water to yield 5 mM final NaOH.
 - If working from 5M NaOH, add **10 µL** of 5M NaOH to **990 µL** of ultrapure water to yield 50 mM NaOH. Mix well and briefly spin down. Take **100 µL** of 50 mM NaOH and combine with **900 µL** of ultrapure water to yield 5 mM NaOH. Mix well and briefly spin down.

Thermal cycler programs

- Preprogram your thermal cycler with these specific programs.
- Use the appropriate programs for the specific Archer assay.
- Verify programming before initiating runs.

	Step	Temperature (°C)	Time (min)
TCR Specific RT Priming	1	65	5
	2	4	Hold

	Step	Temperature (°C)	Time (min)
First Strand cDNA Synthesis	1	50	30
	2	80	20
	3	4	Hold

	Step	Temperature (°C)	Time (min)
Second Strand cDNA Synthesis	1	16	60
	2	75	20
	3	4	Hold

	Step	Temperature (°C)	Time (min)
End Repair (Heated lid off)	1	25	30
	2	4	Hold

	Step	Temperature (°C)	Time (min)
Ligation Step 1	1	37	15
	2	4	Hold

	Step	Temperature (°C)	Time (min)
Ligation Step 2 (Heated lid off)	1	25	15
	2	4	Hold

	Step	Temperature (°C)	Time (min)
Ligation Elution	1	75	10
	2	4	Hold

	Step	Temperature (°C)	Time	Cycles
First PCR	1	95	3 min	1
	2	95	30 sec	24
	3	65	3 min*	
	4	72	3 min	1
	5	4	Hold	1

*See ramp rate guidelines in the “Before getting started” section.

	Step	Temperature (°C)	Time	Cycles
Second PCR	1	95	3 min	1
	2	95	30 sec	8
	3	65	3 min*	
	4	72	3 min	1
	5	4	Hold	1

*See ramp rate guidelines in the “Before getting started” section.

Molecular barcoding, sample indexing, and multiplexed sequencing

Molecule-level barcoding (or unique molecule identifier tagging) and sample-level barcoding (also known as index tagging) are both incorporated during Archer MBC ligation. Molecular barcodes are an integral component of the Archer Analysis software suite (visit archerdx.com/technology-platform/analysis for details). Sample barcodes (i.e., index tags) allow pooled libraries to be sequenced simultaneously thereby enabling maximum sequencing throughput and data demultiplexing during downstream bioinformatics analysis.

Sample multiplexing

- To efficiently utilize the throughput of the MiSeq (or other Illumina sequencing platform) as well as prevent low index diversity within your sequencing run, multiple samples should be sequenced simultaneously. Samples can be identified through a combination of two unique nucleotide sequences (see below for more details), which are subsequently read during the sequencing process. The unique nucleotide sequence is often termed an “index”.

- The Archer Library Preparation Reagents for Illumina use a combination of two indices to distinguish between samples. Index 2 is added during Ligation Step 2 and is embedded in the Archer MBC adapters for Illumina (p5/i5 index). Index 1 is added during the Second PCR and is embedded in Illumina Index 1 Primers (p7/i7) within the Second PCR reaction pellets.
- To maintain appropriate coverage depth, it is recommended that users determine the maximum number of samples that can be run on a flow cell. In general, larger panels with more targets will require higher sequencing coverage depth and should be run with fewer samples.
- Contact archer-tech@idtdna.com to receive tools that can help defining the appropriate number of samples per sequencer and flow cell type.

Barcode diversity

- Illumina sequencers work best when index diversity within a run is high. For example, if eight samples are included in a run, and the user chooses to use only one MBC adapter paired with eight different Index 1 Primers, the run may fail due to low barcode diversity. In this example it is recommended to use eight different Archer MBC adapters paired with eight different Index 1 Primers.
- To facilitate a high index diversity, each tube in the 8-strip of MBC adapters for Illumina as well as Second PCR contains a unique index. When preparing one batch of 8 or less samples, this results in 8 samples with unique dual indices and barcode diversity is not a concern. Only when combining very few samples of different batches, it becomes necessary to pay attention to the index diversity, as the i7 indices (added with Second PCR) are the same in each kit.
- If using more than 48 MBCs, contact archer-tech@idtdna.com for adapter compatibility.

Protocol

Before beginning, review precautions in “Working with Lyophilized Reagents” section for additional information regarding best practices using lyophilized reagents.

Step 1: TCR Specific RT Priming

- Place an appropriate number of **TCR Specific RT Priming** (SA0201) reaction tubes on ice.
- Adjust purified RNA/TNA (20 - 250 ng) samples to a final volume of 20 μ L and transfer to new PCR tubes. Keep input samples on ice.

Component	Reaction Mix
Ultrapure water	20 - X μ l
Purified nucleic acid or RNA	X μ l
<i>Total volume</i>	<i>20μl</i>

- Transfer **20 μ L** reaction mix to the TCR Specific RT Priming reaction tubes.
 - Dissolve, mix and spin down.
 - Return tubes to ice.
- Start the following thermal cycler program, and only transfer reactions to the block once temperature reaches 65°C. Pause the program if necessary.
 - Use a heated lid ($\geq 100^\circ\text{C}$).

TCR Specific RT Priming Incubation Conditions

Step	Temperature ($^\circ\text{C}$)	Time (minutes)
1	65	5
2	4	Hold

- After the program has reached 4°C, briefly spin down reactions and place tubes on ice for **at least 2 minutes**.

Step 2: First Strand cDNA Synthesis

- Place an appropriate number of **First Strand cDNA Synthesis** (SA0002) reaction tubes on ice.

2. Spin down the TCR Specific RT Priming mixture and transfer **20 µL** to the First Strand cDNA Synthesis tubes.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
3. Transfer reactions to a preheated thermal cycler and initiate an incubation using the following program and guidelines:
 - a. Use a heated lid ($\geq 100^{\circ}\text{C}$).

First Strand cDNA Synthesis Incubation Conditions

Step	Temperature ($^{\circ}\text{C}$)	Time (minutes)
1	50	30
2	80	20
3	4	Hold

- b. After the program has reached 4°C , briefly spin down reactions and place on ice.

Step 3: Second Strand cDNA Synthesis

1. Place an appropriate number of **Second Strand cDNA Synthesis** (SA0003) reaction tubes on ice.
2. Add **20 µL** of ultrapure water to each tube containing the **20 µL** of First Strand cDNA Synthesis reaction.
 - a. Pipette up and down to mix.
3. Pipette **40 µL** of each diluted First Strand reaction into the new Second Strand cDNA Synthesis tubes.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
4. Start the following thermal cycler program, and only transfer reactions to the block once temperature reaches 16°C . Pause the program if necessary.
 - a. Use a heated lid ($\geq 100^{\circ}\text{C}$).

Second Strand cDNA Synthesis Incubation Conditions

Step	Temperature (°C)	Time (minutes)
1	16	60
2	75	20
3	4	Hold

- b. After the run has completed, briefly spin down reactions and place on ice.



Safe stopping point: It is okay to stop and store the reactions at -30°C to -10°C.

Step 4: End Repair

1. Place an appropriate number of **End Repair** (SA0204) reaction tubes on ice.
2. Pipette **40 µL** of the Second Strand cDNA Synthesis product into the End Repair tubes.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
3. Transfer reactions to a preheated thermal cycler and initiate an incubation using the following program and guidelines:
 - a. Heated lid off. If the lid is still hot, consider using a different thermal cycler or leaving the lid open.

End Repair Incubation Conditions

Step	Temperature (°C)	Time (minutes)
1	25	30
2	4	Hold

- b. Place samples in the thermal cycler and start the program.
- c. When the run has completed, briefly spin down reactions and place on ice.

Reaction Cleanup after Complete End Repair

Refer to Important Precautions section for guidance on working with AMPure XP beads.

1. Completely resuspend AMPure XP beads by vortexing.
2. Remove tubes from ice and add **2.5X** volume (**100 µL**) of AMPure XP beads to each Complete End Repair reaction.
3. Vortex well or pipette 10 times to mix and visually inspect the color of the sample to ensure a homogenous mixture.
4. Incubate for **5 minutes** at room temperature (20°C to 25°C).
5. Briefly spin down tubes.
6. Place tubes on the magnet for **4 minutes or until beads are fully pelleted** against the tube wall.
7. Without disturbing the bead pellet, use a pipette to remove and discard the supernatant. If the pellet becomes dislodged from the magnet and a portion is drawn into the pipette tip, return contents to the tube and repeat magnet incubation step.
8. Wash beads **two times** with 70% ethanol while still on the magnet. For each wash:
 - a. Add **200 µL** 70% ethanol.
 - b. Incubate for **30 seconds** at room temperature (20°C to 25°C).
 - c. Carefully remove ethanol and discard.
9. After the final wash, use a pipette (≤20 µL capacity) to completely remove visible supernatant residue and allow tubes to dry for **3-5 minutes** at room temperature with open lids. **Take care not to over-dry beads** as this will significantly decrease overall recovery (yield) of nucleic acid.
10. Elute DNA by resuspending beads in **20 µL** 10mM Tris-HCl, pH 8.0.
11. Place tubes back on the magnet for **2 minutes**.

Step 5: Ligation Step 1, 2.0

1. Place an appropriate number of **Ligation Step 1, 2.0** (SA0202) reaction tubes on ice.
2. Transfer **20 µL** of purified cDNA from Reaction Cleanup after End Repair, step 11 into Ligation Step 1, 2.0 tubes. It is acceptable for a small amount of AMPure XP beads to be transferred.

- a. Dissolve, mix and spin down.
- b. Return tubes to ice.
3. Transfer reactions to a preheated thermal cycler and initiate an incubation using the following program and guidelines:
 - a. Use a heated lid ($\geq 100^{\circ}\text{C}$).

Ligation Step 1, 2.0 Incubation Conditions

Step	Temperature ($^{\circ}\text{C}$)	Time (minutes)
1	37	15
2	4	Hold

- b. Place samples in the thermal cycler and start the program.
- c. After the program has reached 4°C , remove tubes from the temperature block, briefly spin down reactions and place on ice.
4. Add **20 μL** 10 mM Tris-HCl, pH 8.0, mix well by pipetting up and down and briefly spin down.

Step 6: MBC Adapter Incorporation

1. Label MBC adapter tubes with the sample index tag letter (A, B, or C) and number (1-48) from the MBC adapters pouch label.
 - a. Use a permanent laboratory marker and orient lid hinges to the back as illustrated:



Important: As this step incorporates the P5 index tag for sample-level tracking, be sure to record which MBC adapter is being used for each sample. Unused tubes must be labeled before returning to the pouch.

2. Place an appropriate number of MBC adapter reaction tubes on ice.
3. Add **40 μL** of the purified cDNA sample from Step 5: Ligation Step 1, 2.0 step 4 to a unique MBC adapter reaction tube.
 - a. Dissolve, mix and spin down.
4. Immediately proceed to Step 7: Ligation Step 2.

Step 7: Ligation Step 2

1. Place an appropriate number of **Ligation Step 2** (SA0197) reaction tubes on ice.
2. Transfer the **entire volume** of each purified cDNA sample from Step 6: MBC Adapter Incorporation to Ligation Step 2 tubes.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
3. Transfer reactions to a thermal cycler and initiate an incubation using the following program and guidelines:
 - a. Heated lid off.

Ligation Step 2 Incubation Conditions

Step	Temperature (°C)	Time (minutes)
1	25	15
2	4	Hold

- b. After the program has reached 4°C, briefly spin down reactions and place on ice.

Safe stopping point: It is okay to stop and store the reactions at -30°C to -10°C.

Reaction Cleanup after Ligation Step 2

Caution: This step uses Ligation Cleanup Beads (SA0210) and Ligation Cleanup Buffer (SA0209) instead of AMPure XP beads and 70% ethanol. Make sure to use Ligation Cleanup Beads and Ligation Cleanup Buffer that come with the kit for this step.

Prepare Ligation Cleanup Beads:

1. Completely resuspend **Ligation Cleanup Beads** by vortexing.
2. For each reaction, pipette **50 µL** of Ligation Cleanup Beads into new 0.2 mL 8-strip tubes.
3. Place tube(s) on the magnet for **1 minute or until the beads are pelleted**.
4. Without disturbing the bead pellet, use a pipette to remove and discard the supernatant. If the pellet becomes dislodged from the magnet and a portion is drawn into the pipette tip, return contents to the tube and repeat magnetic pelleting step.

- Pipette **50 µL of Ligation Cleanup Buffer** into each tube to resuspend beads.

Ligation Cleanup Procedure:



Caution: When vortexing PCR tubes in the subsequent steps, maintain firm pressure on all lids as the contained detergent may allow lids to open.

- Pipette the entire volume of Ligation Step 2 reaction into the tubes with Ligation Cleanup Beads and Buffer.
- Mix samples by vortexing.
- Incubate reactions at room temperature for **5 minutes**.
- Mix samples by vortexing.
- Incubate reactions at room temperature for **5 minutes**.
- Briefly spin down tubes.
- Place tubes on the magnet for **1 minute or until beads are fully pelleted** against the tube wall.
- Carefully pipette off and discard supernatant (**100 µL**) without disturbing the beads.
- Wash beads **two times** with **Ligation Cleanup Buffer**.
 - Resuspend beads in **200 µL** Ligation Cleanup Buffer by vortexing, briefly spinning down, and placing back on magnet for **1 minute**.
 - Once slurry has cleared, discard supernatant.
- Wash beads **once** with **ultrapure water**:
 - Resuspend beads in **200 µL** of ultrapure water by vortexing, briefly spinning down and placing back on magnet.
 - Once slurry has cleared, discard supernatant.
 - Take care to ensure that all supernatant has been removed from tubes.
- Elute DNA from Ligation Cleanup Beads:
 - Resuspend ligation cleanup beads in **36 µL** of 5 mM NaOH.
 - Transfer reactions to a thermal cycler and initiate an incubation using the following program and guidelines:
 - Use a heated lid ($\geq 100^{\circ}\text{C}$).



Ligation Elution Incubation Conditions

Step	Temperature (°C)	Time (minutes)
1	75	10
2	4	Hold

- d. After sample has reached 4°C, briefly spin down and transfer to the magnet.

Step 8: First PCR

1. Place an appropriate number of **First PCR for Illumina-HS** (SA0359) reaction tubes on ice.
 - a. Label tubes by sample number.
2. Pipette **4 µL** of Immunoverse **GSP** into each First PCR tube.
 - a. Spin down and return tubes to ice.
3. Pipette **36 µL** of supernatant from Reaction Cleanup after Ligation Step 2, step 11 into appropriately labeled First PCR tubes.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
4. Transfer reactions to a preheated thermal cycler and immediately initiate the program below.
 - a. Use a heated lid (≥100°C).

First PCR Incubation Conditions

Step	Temperature (°C)	Time	Cycles
1	95	3 min	1
2	95	30 sec	24
3	65	3 min*	
4	72	3 min	1
5	4	Hold	1

*See ramp rate guidelines in the “Before getting started” section.

- b. After the program has reached 4°C, briefly spin down reactions and place on ice. It is also acceptable to leave tubes in the thermal cycler at 4°C overnight.

Reaction Cleanup after First PCR

Refer to Important Precautions section for guidance on working with AMPure XP beads.

1. Completely resuspend AMPure XP beads by vortexing.
2. Add **1.2X** volume (**48 µL**) of AMPure XP beads to each reaction.
3. Vortex well or pipette 10 times to mix and visually inspect the color of the sample to ensure a homogenous mixture.
4. Incubate for **5 minutes** at room temperature (20°C to 25°C).
5. Briefly spin down tubes.
6. Place tubes on the magnet for **4 minutes or until beads are fully pelleted** against the tube wall.
7. Without disturbing the bead pellet, use a pipette to remove and discard the supernatant. If the pellet becomes dislodged from the magnet and a portion is drawn into the pipette tip, return contents to the tube and repeat magnet incubation step.
8. Wash beads **two times** with 70% ethanol while still on the magnet. For each wash:
 - a. Add **200 µL** 70% ethanol
 - b. Incubate for **30 seconds** at room temperature (20°C to 25°C)
 - c. Carefully remove ethanol and discard
9. After the final wash, use a pipette (≤20 µL capacity) to completely remove visible supernatant residue and allow tubes to dry for **3-5 minutes** at room temperature with open lids. **Take care not to over-dry beads** as this will significantly decrease overall recovery (yield) of nucleic acid.
10. Elute DNA by resuspending beads in **44 µL** of 10 mM Tris-HCl, pH 8.0.
11. Place tubes back on the magnet for **2 minutes**.
12. Transfer **42 µL** of purified eluate to a new 0.2mL PCR tube and store reactions as indicated below or proceed directly to Step 9: Second PCR.



Safe stopping point: It is okay to stop and store the reactions at -30°C to -10°C (or leave in thermal cycler on 4°C hold for no more than 24 hours).

Step 9: Second PCR

1. Place an appropriate number of **Second PCR Reactions 1 thru 8 for Illumina-HS (SA0360)** reaction tubes on ice.



Important: The Index 1 (P7) index tag is incorporated during this step.

- a. Use a permanent marker to label the tubes 1 to 8 from left to right as shown below. (Review the Molecular Barcoding, Sample Indexing & Multiplexed Sequencing section above).



- b. Unused tubes must be labeled before returning to the pouch.
2. Pipette **40 µL** of purified First PCR reaction into each Second PCR tube on ice.
 - a. Dissolve, mix and spin down.
 - b. Return tubes to ice.
3. Transfer reactions to a preheated thermal cycler and immediately initiate the program below.
 - a. Use a heated lid (≥100°C).



Second PCR Incubation Conditions

Step	Temperature (°C)	Time	Cycles
1	95	3 min	1
2	95	30 sec	8
3	65	3 min*	
4	72	3 min	1
5	4	Hold	1

*See ramp rate guidelines in the “Before getting started” section.

- b. After the program has reached 4°C, briefly spin down reactions and place on ice. It is also acceptable to leave tubes in the thermal cycler at 4°C overnight.

Reaction Cleanup after Second PCR

Refer to Important Precautions section for guidance on working with AMPure XP beads.

1. Completely resuspend AMPure XP beads by vortexing.
2. Add **1.2X** volume (**48 µL**) of AMPure XP beads to each Second PCR reaction.
3. Vortex well or pipette 10 times to mix and visually inspect the color of the sample to ensure a homogenous mixture.

4. Incubate for **5 minutes** at room temperature (20°C to 25°C).
5. Briefly spin down tubes.
6. Place tubes on the magnet for **4 minutes or until beads are fully pelleted** against the tube wall.
7. Without disturbing the bead pellet, use a pipette to remove and discard the supernatant. If the pellet becomes dislodged from the magnet and a portion is drawn into the pipette tip, return contents to the tube and repeat magnet incubation step.
8. Wash beads **two times** with 70% ethanol while still on the magnet. For each wash:
 - a. Add **200 µL** 70% ethanol.
 - b. Incubate for **30 seconds** at room temperature (20°C to 25°C).
 - c. Carefully remove ethanol and discard.
9. After the final wash, use a pipette (≤20 µL capacity) to completely remove visible supernatant residue and allow tubes to dry for **3-5 minutes** at room temperature with open lids. **Take care not to over-dry beads** as this will significantly decrease overall recovery (yield) of nucleic acid.
10. Elute DNA by resuspending beads in **20 µL** 10 mM Tris-HCl, pH 8.0.
11. Place tubes back on the magnet for **2 minutes**.
12. Transfer **18 µL** of the purified solution to a new 0.2mL PCR tube. Be sure to avoid transferring beads to the fresh tube.
13. Stop or proceed directly to Quantify, Normalize and Sequence.



Safe stopping point: It is okay to stop and store the reactions at -30°C to -10°C.

Quantify, Normalize, and Sequence

Refer to the Quantify, Normalize, and sequence Protocol for Illumina (RA-DOC-054) for information on quantifying, normalizing and sequencing Archer AMP libraries.

Data analysis

Archer AMP libraries should be demultiplexed to FASTQs prior to analysis. Analyze data with Archer Analysis using either a local software installation or Archer Unlimited. Visit our website or contact archer-tech@idtdna.com for more information.

Immunoverse assays require a one-time upload of a GTF file (a text file, in GTF format, which directs the software on how to analyze data from the panel). This file can be obtained by contacting archer-tech@idtdna.com.

©2023 IntegratedDNA Technologies, Inc. All rights reserved. Unless otherwise specified, all product names, service names, and logos appearing in this product insert are trademarks owned by or licensed to Archer, its subsidiaries, or its affiliates. No use of any Archer trademark, trade name, or trade dress in this product insert may be made without the prior written authorization of Archer, except to identify Archer's products or services.

This product or service is licensed under one or more of the following U.S. Patents: 8,835,358; 9,290,808; 9,290,809; 9,315,857; 9,708,659; and 9,816,137 owned by BD, and is licensed solely for the use described in the associated product literature. No other rights, implied or otherwise, are granted to purchaser hereunder. Purchaser agrees, by way of example and not limitation, not to use this product to trace back the origin of a nucleic acid to an individual cell as a discrete entity (e.g. single cell analysis).]