

## Product Insert VariantPlex™ CFTR Panel

### VariantPlex CFTR

#### **Description**

The VariantPlex CFTR panel is a balanced pool of gene-specific primer (GSP) oligonucleotides that is optimized for use with VariantPlex reagents and molecular barcode (MBC) adapters to produce targeted NGS libraries. This product insert should be used in conjunction with VariantPlex standard protocol (RA-DOC-057).

VariantPlex CFTR contains **105** GSPs with coverage of all exons of the Cystic Fibrosis Conductance Regulator (CFTR) gene, including 5' and 3' untranslated regions (UTRs) and select intronic variants.

Description	Part number	Storage	
VariantPlex CFTR GSP1, 8 reactions	SA0079-8-1	−20°C ± 10°C	
VariantPlex CFTR GSP2, 8 reactions	SA0079-8-2	20 C ± 10 C	

### **Recommended PCR cycling**

	Step	Temperature (°C)	Time	Cycles
First PCR reaction	1	95	3 min	1
	2	95	30 sec	14
	3	65	5 min (100% ramp rate)	
	4	72	3 min	1
	5	4	Hold	1

2425 55<sup>th</sup> Street, Boulder, CO 80301 | www.archerdx.com | archer-tech@idtdna.com



### Product Insert VariantPlex™ CFTR Panel

#### Recommended PCR cycling (cont.)

	Step	Temperature (°C)	Time	Cycles
Second PCR reaction	1	95	3 min	1
	2	95	30 sec	
	3	65	5 min (100% ramp rate)	22†
	4	72	3 min	1
	5	4	Hold	1

<sup>†</sup>The number of PCR2 cycles may be decreased if you regularly experience library yields greater than 200 nM.

#### Recommended reads and multiplexing

VariantPlex CFTR libraries should be sequenced to a minimum of **500,000 reads**. Starting read depth recommendations for standard profiling may be adjusted based on user needs.

#### **Archer™ Analysis settings**

Sequencing data should be processed using Archer Analysis (v7.0 Version, or greater). The VariantPlex CFTR panel requires selection of the *SNV/Indel and Structural Variation* pipeline, found under the *DNA* Input Type (see the Archer Analysis User Guide for more details on setting up your analysis). Selection of the DNA Target Coverage pipeline is optional.

Processing of VariantPlex CFTR libraries requires a one-time upload of the Panel GTF. When performing DNA Target Coverage analysis, users must also select a Region of Interest BED file. Users may optionally add a Targeted Mutations VCF file for targeted SNV/Indel detection. Files can be obtained by contacting <a href="mailto:archer-tech@idtdna.com">archer-tech@idtdna.com</a>

2425 55th Street, Boulder, CO 80301 | www.archerdx.com | archer-tech@idtdna.com



# Product Insert VariantPlex™ CFTR Panel

#### Limitations of use

For research use only. Not for use in diagnostic procedures. Unless otherwise agreed to in writing, IDT does not intend these products to be used in clinical applications and does not warrant their fitness or suitability for any clinical diagnostic use. Purchaser is solely responsible for all decisions regarding the use of these products and any associated regulatory or legal obligations.

Safety data sheets pertaining to this product are available upon request.

This product or service is licensed under one or more of the following U.S. Patents: 8,835,358; 9,290,808; 9,290,809; 9,315,857; 9,708,659; and 9,816137 owned by BD and is licensed solely for the use described in the associated product literature. No other rights, implied or otherwise, are granted to purchaser hereunder. Purchaser agrees, by way of example and not limitation, not to use this product to trace back the origin of a nucleic acid to an individual cell as a discrete entity (e.g., single cell analysis).

© 2023 Integrated DNA Technologies, Inc. All rights reserved. FusionPlex, VariantPlex, LiquidPlex, Immunoverse, Archer Analysis, and Archer Assay Marketplace are trademarks of Integrated DNA Technologies, Inc. All other marks are the property of their respective owners. For specific trademark and licensing information, see <a href="https://www.idtdna.com/trademarks">www.idtdna.com/trademarks</a>.